

Zero-field remote detection of NMR with a microfabricated atomic magnetometer

M. P. Ledbetter*, I. M. Savukov*, D. Budker*[†], V. Shah[‡], S. Knappe[‡], J. Kitching[‡], D. J. Michalak[§], S. Xu[§], and A. Pines^{§†}

*Department of Physics, University of California, Berkeley, CA 94720-7300; [†]Nuclear Science Division, Lawrence Berkeley National Laboratory, Berkeley, CA 94720; [‡]Time and Frequency Division, National Institute of Standards and Technology, 325 Broadway, Boulder, CO 80305; and [§]Department of Chemistry, University of California, Berkeley, CA 94720-7300

Contributed by A. Pines, December 18, 2007 (sent for review October 3, 2007)

We demonstrate remote detection of nuclear magnetic resonance (NMR) with a microchip sensor consisting of a microfluidic channel and a microfabricated vapor cell (the heart of an atomic magnetometer). Detection occurs at zero magnetic field, which allows operation of the magnetometer in the spin-exchange relaxation-free (SERF) regime and increases the proximity of sensor and sample by eliminating the need for a solenoid to create a leading field. We achieve pulsed NMR linewidths of 26 Hz, limited, we believe, by the residence time and flow dispersion in the encoding region. In a fully optimized system, we estimate that for 1 s of integration, 7×10^{13} protons in a volume of 1 mm^3 , prepolarized in a 10-kG field, can be detected with a signal-to-noise ratio of ≈ 3 . This level of sensitivity is competitive with that demonstrated by microcoils in 100-kG magnetic fields, without requiring superconducting magnets.

microfluidics | signal-to-noise ratio | mass-limited sample

Remote detection of nuclear magnetic resonance (NMR) (1), in which polarization, encoding or evolution, and detection are spatially separated, has recently attracted considerable attention in the context of magnetic resonance imaging (2), microfluidic flow profiling (3, 4), and spin-labeling (5). Detection can be performed with superconducting quantum interference devices (SQUIDs), inductively at high field as in refs. 3–5 or with atomic magnetometers as in ref. 2. To most efficiently detect the flux from the nuclear sample, it is typically necessary to match the physical dimensions of the sensor and the sample. Thus, small, sensitive detectors of magnetic flux reduce the detection volume, thereby reducing the quantity of analyte. Microfabricated atomic magnetometers (6) with sensor dimensions on the order of 1 mm operating in the spin-exchange relaxation-free (SERF) regime (8) have recently demonstrated sensitivities of $\approx 0.7 \text{ nG}/\sqrt{\text{Hz}}$ (7), with projected theoretical sensitivities several orders of magnitude higher. (In this article, we use Gaussian units; $1 \text{ nG} = 100 \text{ fT}$.)

In this work, we demonstrate remote detection of pulsed and continuous-wave (CW) NMR with a compact sensor assembly consisting of an alkali vapor cell and microfluidic channel, fabricated with lithographic patterning and etching of silicon. We realize pulsed NMR linewidths of $\approx 26 \text{ Hz}$, limited, we believe, by residence time and flow dispersion in the encoding region. Estimates of the fundamental sensitivity limit for an optimized system, assuming a modest 10-kG prepolarizing field, indicate detection limits competitive with those demonstrated by microcoils in superconducting magnets (9–14). Hence, the technique described here offers a promising solution to NMR of mass-limited samples—for example, in the screening of new drugs—without requiring superconducting magnets.

The atomic magnetometer operates in the SERF regime (achieved when the Larmor precession frequency is small compared with the spin exchange rate), currently the most sensitive technique in atomic magnetometry. Optical pumping and probing of the alkali vapor are accomplished with a single laser beam (7). In addition to integration of sensor and microfluidic channel

on a single chip, a key feature distinguishing the present work from previous applications of atomic magnetometers to the detection of NMR (15, 16) or MRI (2) is that the detection region (both magnetometer and nuclear sample) is at zero magnetic field. This eliminates the need for a solenoid around the detection region (along with the associated noise) and increases the proximity of sensor and sample. A secondary advantage of having both sensor and sample at zero field is that it is the only point at which the Zeeman resonance frequencies of both alkali and nuclear spins coincide, yielding sensitivity to all three components of the nuclear magnetization. This may prove important for the development of new algorithms for efficient remote detection.

The experimental setup is shown in Fig. 1. Tap water flows through 800- μm -inner-diameter Teflon tubes from a reservoir in a prepolarizing field to an encoding region and finally through the microchip, which is housed inside a four-layer set of magnetic shields. The prepolarizing field is provided by a 7-kG permanent magnet, and the volume of the reservoir is $\approx 10 \text{ cm}^3$, large enough that the water spends several longitudinal relaxation times (T_1) in the prepolarization field. The encoding region consists of a bulb $\approx 4 \text{ mm}$ in diameter and a Helmholtz coil used to apply audio frequency (AF) pulses. An anti-Helmholtz coil was used to shim longitudinal gradients of the ambient magnetic field in our laboratory ($\approx 250 \text{ mG}$) to a uniformity of $\approx 1 \text{ mG}$ over a 5-cm baseline. For the present set of measurements, the flow rate was $\approx 1.5 \text{ ml/s}$, so that the time required to fill and empty the bulb was approximately $\tau_{\text{res}} = 22 \text{ ms}$ and the distance from the encoding region to the detection region was $\approx 50 \text{ cm}$.

The sensor chip consists of a vapor cell (the atomic magnetometer) and channel constructed by anodically bonding glass to a 1-mm-thick etched Si substrate. The transverse dimensions of the vapor cell and fluid channel are both $\approx 2 \times 3 \text{ mm}$. The fabrication techniques are described in more detail in ref. 17. The vapor cell contains 5,000 torr of N_2 buffer gas and Cs, chosen primarily because its saturated vapor pressure is higher than that of K or Rb, allowing operation of the sensor at lower temperatures. The vapor cell is sandwiched between two indium tin-oxide (ITO) resistive heaters. The heaters, powered by DC current (AC current caused several heaters to crack for unknown reasons), saturated the magnetometer, necessitating cycling of the heaters on and off. When the heaters were off, considerable (repeatable) drift in magnetometer signal occurred due to temperature drift; hence, alternate measurement periods were used for the acquisition of the NMR signal and recording of background drift. To optimize the magnetometer, the current in the heaters was adjusted until the cell comprised approximately one absorption length. We did not measure the temperature

Author contributions: M.P.L., D.B., J.K., and A.P. designed research; M.P.L., I.M.S., V.S., D.J.M., and S.X. performed research; V.S. and S.K. contributed new reagents/analytic tools; M.P.L. analyzed data; and M.P.L., D.B., S.K., J.K., D.J.M., and S.X. wrote the paper.

The authors declare no conflict of interest.

[†]To whom correspondence should be addressed. E-mail: pines@berkeley.edu.

© 2008 by The National Academy of Sciences of the USA




Fig. 4. Free induction decay observed by applying two $\pi/2$ pulses separated by a delay. Overlaying the data is a fit to a numerical model that includes the finite length of the pulses and counter-rotating components of the AF, yielding $T_2 = 6$ ms.

this initial fit, we then fit the signal following a pulse sequence to $af_\pi(t)$ with only the signal amplitude a as the free parameter. The solid line overlaying the data in Fig. 4 is a fit to a numerical model that includes the effects of finite pulse length and counter-rotating components of the applied AF magnetic field. Parameters in the fit are the initial signal amplitude S_0 , the Larmor precession frequency ν_0 , and the transverse relaxation time T_2 . The phase of the signal is determined by the detuning of the AF from the Larmor precession frequency. We find $\nu_0 = 1,157$ Hz and $T_2 = 6.0$ ms corresponding to $\Delta\nu \equiv 1/(2\pi T_2) = 26$ Hz. The sampling interval for these measurements was 2 ms, yielding a Nyquist frequency of 250 Hz, not sufficient to fully resolve Larmor precession in the neighborhood of 1 kHz. However, from measurements of the laboratory field and from CW NMR measurements (see below), the resonant frequency can be constrained to $\approx 1,100$ – $1,200$ Hz, and hence we are confident in the values of the parameters obtained from the fit.

We estimate the spread in Larmor frequencies due to measured magnetic field gradients to be approximately $\Delta\nu_g = 0.3$ Hz, far too small to account for the observed width. Broadening due to the finite residence time in the encoding region is difficult to calculate because of flow dispersion. A minimum value is $\approx 1/(2\pi\tau_{\text{res}}) = 7$ Hz, approximately a factor of 4 smaller than the measured width, illustrating the importance of minimizing dispersion for high-resolution remotely detected NMR. We anticipate that significantly narrower lines will result from better control over dispersion.

NMR can also be observed by continuous application of AF magnetic field in the encoding region. The resulting signal as a function of frequency is shown in Fig. 5 for relatively weak AF. Fig. 5 *Inset* shows the dependence of the signal on the amplitude of the AF for $\nu - \nu_0 \approx 0$. The signal saturates for large AF amplitude because of the broad range of residence times in the encoding region due to dispersion. To minimize AF power broadening, the data in Fig. 5 were taken for the smallest applied amplitude of AF shown in Fig. 5 *Inset*, corresponding to $B_1 \approx 4$ mG in the rotating frame. Overlaying the data is a fit to an absorptive Lorentzian with half-width at half-maximum $\Delta\nu = 42 \pm 3$ Hz. We suspect that the difference in linewidths obtained by pulsed and CW NMR is due to AF power broadening in conjunction with dispersion-induced variation in residence times in the encoding region.

We now turn to an optimization of remote detection of NMR with the presently considered device, beginning with an estimate of the theoretical sensitivity of the atomic magnetometer to magnetic fields created by the polarized nuclei. The sensitivity of




Fig. 5. Magnetometer signal for continuous application of weak AF in encoding region as a function of frequency. The solid line overlaying the data is a fit to an absorptive Lorentzian, resulting in a half-width at half-maximum $\Delta\nu = 43$ Hz. (*Inset*) Signal as a function of amplitude of the AF, tuned to resonance (units on the vertical axis are the same as those in the main figure). The data in the main figure were obtained at the minimum amplitude value of applied AF shown in *Inset*.

the magnetometer is fundamentally limited by spin-projection noise (see, e.g., ref. 19) $\delta B \approx \hbar/(g_s\mu_B\sqrt{nV\tau/\Gamma})$, where $g_s \approx 2$, μ_B is the Bohr magneton, V is the volume of the sensor, Γ is the alkali relaxation rate, and τ is the measurement time. For sufficiently high alkali densities, binary alkali–alkali spin-destruction collisions dominate the relaxation rate, and spin-projection noise approaches an asymptote given by

$$\delta B \approx \frac{\hbar}{g_s\mu_B} \sqrt{\frac{\bar{v}\sigma_{\text{sd}}}{V\tau}}, \quad [1]$$

where \bar{v} is the mean relative velocity of colliding alkali atoms and σ_{sd} is the alkali–alkali spin-destruction cross-section. Prefactors of order unity in the right-hand side of Eq. 1 depend on the particulars of the pumping and probing scheme. If the sensor and sample are the same size and are separated by a small distance, the magnetic field the sensor experiences is approximately equal to the magnetization M . The thermal magnetization of protons in a prepolarizing field B_p is $M = \mu_p^2(N_p/V) B_p/kT$, where μ_p is the proton magnetic moment and N_p is the total number of protons in volume V . Hence, from Eq. 1, a signal-to-noise ratio of 3 or greater requires a minimum number of protons

$$\delta N_p = 3 \times \frac{\hbar kT}{g_s\mu_B\mu_p^2 B_p} \sqrt{\frac{\bar{v}\sigma_{\text{sd}}V}{\tau}}. \quad [2]$$

To obtain concrete numbers, we assume a model device with sensor and sample both of volume $V = 1$ mm³ and use spin-destruction cross-sections appropriate for Rb [$\sigma_{\text{sd}} = 9 \times 10^{-18}$ cm² (20), approximately a factor of 20 smaller than that of Cs (21)]. The contribution to relaxation from collisions with the cell walls and buffer gas atoms can be significant (we estimate that the contribution to the relaxation rate from wall and buffer gas collisions reaches a minimum value of $\approx 3,000$ s⁻¹ with $\approx 4,000$ torr of N₂ buffer gas in a 1-mm³ Rb vapor cell), and hence, to reach the asymptotic limit of magnetometric sensitivity given by Eq. 1, Rb number densities of $\approx 7 \times 10^{15}$ cm⁻³ are required, corresponding to operating temperatures of $\approx 250^\circ\text{C}$. (For such high alkali densities, the cell under consideration comprises ≈ 40 absorption lengths, in which case it may be necessary to monitor optical rotation of a separate probe beam tuned far off resonance.) Under these conditions, with a prepolarizing field of $B_p = 10$ kG, Eq. 2 gives a detection limit of approximately $\delta N_p \approx$

7×10^{13} protons or ≈ 120 pmol (corresponding to a concentration of $120 \mu\text{M}$) for $\tau = 1$ s of integration. This is competitive with the detection limit demonstrated by microcoils in high magnetic fields (see, e.g., ref. 14, in which $\approx 5 \times 10^{13}$ protons were detected with a signal-to-noise ratio of 1 in 1 s of integration in a 383-MHz superconducting magnet).

The magnetometric sensitivity required to reach the fundamental detection limit for the conditions stated above is $\approx 1.7 \text{ pG}/\sqrt{\text{Hz}}$, approximately 3 orders of magnitude better than that achieved in this work. Realizing this level of sensitivity is admittedly quite challenging and will likely require advanced techniques in magnetic shielding (see, e.g., ref. 22) and ultra-low-noise lasers. Monitoring optical rotation of a separate, far off resonant probe beam (necessary to efficiently probe an optically thick vapor cell such as that considered above) has the additional advantage that noise due to laser intensity fluctuations can be canceled out. By virtue of the \sqrt{V} dependence in Eq. 2, smaller magnetometers operating at higher temperatures could lead to further improvements in sensitivity. However, at sufficiently high temperatures, alkali atoms begin to react with the glass, which will place a lower bound on the volumes over which Eq. 2 is valid.

Mapping of transverse components of the magnetization onto the longitudinal component, as in the present work, results in an interesting consequence for signal acquisition when integration times are long compared with T_1 . Consider a remote phase-encoding experiment in which one wishes to collect N data points of a free induction decay using a detector with sensitivity ρ_r (with units $\text{G}/\sqrt{\text{Hz}}$). We assume that a stop-flow arrangement is used so that fluid can reside in the detection region for as long as we wish but can be transferred from the encoding region to the detection region in a time short compared with T_1 . The signal in the detection region following a phase encoding sequence that results in a longitudinal component of magnetization M_{\parallel} is then $M_{\parallel}e^{-t/T_1}$. It is straightforward to show that the optimal signal-to-noise ratio is obtained for a measurement time equal to $t = 1.25 T_1$ resulting in an uncertainty of the phase encoded amplitude $\sigma_r^{\text{min}} = 2.21\rho_r/\sqrt{T_1}$. After acquiring a single point, the fluid must be repolarized so that the time required to measure a single point is κT_1 , where κ is a dimensionless parameter of order unity, yielding a total experiment time $T_{\text{exp}} = N\kappa T_1$.

Now consider a typical inductively detected direct experiment, in which one collects N data points in a single transient lasting T_2 . The detector must operate with a bandwidth $BW = 1/(2\Delta t)$, where $\Delta t = T_2/N$ is the sampling interval. Hence, the uncertainty of each point is $\rho_d\sqrt{N}/(2T_2)$, where ρ_d is the sensitivity of the direct detector. The repetition time (limited by T_1) may be taken to be the same in the direct and remote modes, κT_1 , so that in an experiment lasting the same amount of time as in the remote case, a total number of transients $N_t = T_{\text{exp}}/\kappa T_1 = N$ may be collected. After averaging N_t transients, the uncertainty of each point in the directly detected FID is reduced by $1/\sqrt{N_t}$, so that $\sigma_d = \rho_d/\sqrt{2T_2}$. Comparing the uncertainty of each point in the remote and direct modes, we find $\sigma_r/\sigma_d = 1.5 \rho_r T_2/(\rho_d T_1)$. Hence, for detectors of comparable sensitivity, an experiment that

detects the longitudinal component of the magnetization is more sensitive by a factor $1.5\sqrt{T_1/T_2}$.

The present technique appears to have several limitations. First, the high temperature estimated for optimal operation of the magnetometer may be prohibitive for analysis of organic compounds that become unstable at high temperature. The most obvious solution to this problem is efficient thermal isolation of the vapor cell and microfluidic channel. For example, the vapor cell could be mounted in close proximity to the microfluidic channel via low thermal conductivity polyimide tethers as in ref. 23, where ≈ 9 mW of heating power was required to heat similarly sized vapor cells to 95°C in vacuum. Another possible solution is the use of antirelaxation wall coatings in the vapor cell. This would allow the use of lower buffer gas pressures, so that alkali-alkali spin-destruction collisions would dominate at lower temperatures, thereby lowering the temperature at which the asymptotic limit of magnetometric sensitivity (Eq. 1) would be reached. The best coating presently available is paraffin, allowing up to 10,000 bounces before depolarizing atoms; however, paraffin typically does not survive temperatures beyond 70°C . A promising alternative coating is octadecyltrichlorosilane, which has been shown to operate at temperatures of up to 120°C (24).

Finally, in the present work, encoding was performed in a field of ≈ 250 mG. This value of magnetic field allows access to scalar (J) couplings; however, chemical shifts, which are typically on the order of several parts per million for ^1H , would be difficult to observe. Future work will likely explore the possibility of measuring chemical shifts by employing reasonably homogeneous permanent magnets and spatially tailored RF fields with appropriate pulse sequences to counteract the effects of inhomogeneous magnetic fields, as has been demonstrated in refs. 25 and 26.

In conclusion, we have demonstrated remote detection of both pulsed and CW NMR with a SERF magnetometer/microfluidic channel integrated into a single microfabricated device. We realized pulsed NMR linewidths of ≈ 26 Hz, limited, we believe, by the residence time and flow dispersion in the encoding region. Measurements were performed at zero field, allowing operation in the SERF regime and eliminating the need for a solenoid surrounding the sample, increasing the proximity of sensor and sample. Estimates of the fundamental detection limit indicate that, for an integration time of 1 s and a relatively modest prepolarizing field of 10 kG, $\approx 7 \times 10^{13}$ protons can be detected in a volume of 1 mm^3 with a signal-to-noise ratio of 3. With fast algorithms for remote detection and a recirculating pump to minimize the total volume of analyte, the technique presented here offers a promising alternative to conventional detection of NMR at high field.

ACKNOWLEDGMENTS. We greatly appreciate stimulating discussions with L.-S. Bouchard. This work was supported by Office of Naval Research–Multidisciplinary University Research Initiative Grant FD-N00014-05-1-0406; the Director, Office of Science, Office of Basic Energy Sciences, Nuclear Science Divisions, of the U.S. Department of Energy under Contract DE-AC03-76SF00098; a CalSpace Minigrant; and the Microsystems Technology Office of the Defense Advanced Research Projects Agency.

- Moulé AJ, et al. (2003) Amplification of xenon NMR, MRI by remote detection. *Proc Natl Acad Sci USA* 100:9122–9127.
- Xu S, et al. (2006) Magnetic resonance imaging with an optical atomic magnetometer. *Proc Natl Acad Sci USA* 103:12668–12671.
- Granwehr J, et al. (2007) Dispersion measurements using time-of-flight remote detection MRI. *Magn Reson Imaging* 25:449–452.
- Harel E, Hilty C, Koen K, McDonnell EE, Pines A (2007) Time-of-flight flow imaging of two-component flow inside a microfluidic chip. *Phys Rev Lett* 98:017601.
- Anwar MS, et al. (2007) Spin coherence transfer in chemical transformations monitored by remote detection NMR. *Anal Chem* 79:2806–2811.
- Schwindt PDD, et al. (2004) Chip-scale atomic magnetometer. *Appl Phys Lett* 85:6409–6411.
- Shah V, Knappe S, Schwindt PDD, Kitching J (2007) Subpicotesla atomic magnetometry with a microfabricated vapour cell. *Nat Photonics* 1:649–652.
- Kominis IK, Kornack TW, Allred JC, Romalis MV (2003) A subfemtotesla multichannel atomic magnetometer. *Nature* 422:596–599.
- Olson DL, Peck TL, Webb AG, Margin RL, Sweedler JV (1995) High-resolution microcoil ^1H -NMR for mass-limited, nanoliter-volume samples. *Science* 270:1967–1970.
- Olson DL, Lacey ME, Sweedler J (1998) The nanoliter niche—NMR detection for trace analysis and capillary separations. *Anal Chem* 70:645–650.
- Webb A (2007) Nuclear magnetic resonance of mass-limited samples using small RF coils. *Anal Bioanal Chem* 388:525–528.
- Djukovic D, Liu S, Henry I, Tobias B, Raftery D (2006) Signal enhancement in HPLC/microcoil NMR using automated column trapping. *Anal Chem* 78:7154–7160.
- Minard KR, Wind RA (2002) Picoliter ^1H NMR spectroscopy. *J Magn Reson* 154:336–343.
- Ciobanu L, Seeber DA, Pennington CH (2002) 3D MR microscopy with resolution $3.7 \mu\text{m}$ by $3.3 \mu\text{m}$ by $3.3 \mu\text{m}$. *J Magn Reson* 158:178–182.
- Savukov IM, Romalis MV (2005) NMR detection with an atomic magnetometer. *Phys Rev Lett* 94:123001.

16. Savukov IM, Seltzer SJ, Romalis MV (2007) Detection of NMR signals with a radio-frequency atomic magnetometer. *J Magn Reson* 185:214–220.
17. Knappe S, et al. (2005) Atomic vapor cells for chip-scale atomic clocks with improved long-term frequency stability. *Opt Lett* 30:2351–2353.
18. Seltzer SJ, Romalis MV (2004) Unshielded three-axis vector operation of a spin-exchange-relaxation-free atomic magnetometer. *Appl Phys Lett* 85:4804–4806.
19. Budker D, Romalis MV (2007) Optical magnetometry. *Nat Phys* 3:227–234.
20. Baranga AB, et al. (1998) Polarization of ^3He by spin exchange with optically pumped Rb and K vapors. *Phys Rev Lett* 80:2801–2804.
21. Beverini N, Minguzzi P, Strumia F (1971) Foreign-gas-induced cesium hyperfine relaxation. *Phys Rev A* 4:550.
22. Kornack TW, Smullin SJ, Lee S-K, Romalis MV (2007) A low-noise ferrite magnetic shield. *Appl Phys Lett* 90:223501.
23. Lutwak R, et al. (2004) *The Chip-Scale Atomic Clock—Low-Power Physics Package*, Proceedings of the 36th Annual Precise Time and Time Interval (PTTI) Systems and Applications Meeting, ed Senior K (National Technical Information Service, Springfield, VA), pp 339–354.
24. Seltzer SJ, Meares PJ, Romalis MV (2007) Synchronous optical pumping of quantum revival beats for atomic magnetometry. *Phys Rev A* 75:051407.
25. Meriles CA, Sakellariou D, Heise H, Moule AJ, Pines A (2001) Approach to high-resolution ex situ NMR spectroscopy. *Science* 293:82–85.
26. Grunin L, Blümich B (2004) Resolving chemical shift spectra with a low-field NMR relaxometer. *Chem Phys Lett* 397:306–308.